

DIGITAL INSULATION/CONTINUITY TESTER KEW 3551 / 3552 / 3552BT



LED light & Display backlight

Facilitate working at dimly illuminated location.

Illuminance sensor turns the LCD backlight and LED spot light ON/OFF.







Diagnostic Insulation Tests Polarization Index



Insulation resistance value 1 min. after start	
Pl 4.0 or more 4.0 to 2.0 2.0 to 1.0 1.0 or loss	11-
FI 4.0 01 11016 4.0 to 2.0 2.0 to 1.0 1.0 01 less	PI
Criteria Best Good Warning Bad	Criteria



Dielectric Absorption Ratio							
DAR= —	Insulation resistance value 1 min. after start						
DAN-	Insulation resistance value 15 sec. after start						
DAR	1.4 or more	1.25 to 1.0	1.0 or less				
Criteria	Best	Good	Bad				

Memory/data transfer function (available on 3552/3552BT)

Internal memory up to 1000 measurements can be transferred to a PC by the optional adapter 8212-USB.



*Please download the software from our website

Provide maximum versatility with the following functions.

- Insulation resistance measurement (6 ranges)
- AC/DC voltage measurement
- Continuity test @200mA, Low-resistance



Bluetooth® communication (3552BT)

Transfer measured data directly to the compatible tablet or Smart phone devices.

Save time - easy to create reports.

Available on download site for free.



KEW 3551 / 3552 / 3552BT Specification

nsulation resistance							
Test voltage	50V	100V	125V	250V	500V	1000V	
Measuring range (Auto-ranging)	4.000/40.00/ 100.0MΩ	4.000/40.00/ 200.0MΩ	4.000/40.00/ 250.0MΩ	4.000/40.00/ 400.0/500.0MΩ	4.000/40.00/ 400.0/2000MΩ /20GΩ ^{*1}	4.000/40.00/ 400.0/4000ΜΩ /40GΩ ^{*1}	
Mid-scale value	2MΩ $5MΩ$ $10MΩ$ $100MΩ$ $200MΩ$						
First effective measuring range	0.100 to 10.00MΩ						
Accuracy	±2%rdg±2dgt						
Second effective measuring range	0.050 to 0.099ΜΩ						
	10.01 to 100.0MΩ 20.01 to 200.0MΩ 25.01 to 250.0MΩ 50.1 to 500MΩ 501 to 2000MΩ 1001 to 4000MΩ						
Accuracy	±5%rdg						
Rated current							
	@0.05MΩ	@0.1MΩ	@0.125MΩ	@0.25MΩ	@0.5MΩ	@1MΩ	
Output short circuit current	1.5mA max						
2/Continuity ^{*2}							
Resistance range	40.00/400.0/4000Ω						
Accuracy	±2.5%rdg±8dgt						
Open-circuit voltage	5V(4 to 6.9V)						
Measuring current	200mA or more						
/oltage							
Range	2.0 to 600V AC(45 to 65Hz)±2.0 to ±600V DC						
Accuracy	±1%rdg±4dgt						
General							
Applicable standards	IEC 61010 CAT III 600V / CAT IV 300V						
	IEC 61010-2-034, IEC 61557-1,2,4, IEC 61326-1,2-2, IEC 60529(IP40)						
Communication interface	USB *1, Bluetooth®5.0 *2						
Dimension / Weight	97(L)x156(W)x46(D)mm / Approx. 490g(including battery)						
Power source	LR6/R6(AA)(1.5V) x 4						
Accessories	7260(Test lead with remote control switch), 7261A(Test lead with alligator clip), 8017A(Extension prod long), 9173(Carrying case), 9121(Shoulder strap), Batteries, Instruction manua						
Optional accessories	7243A(L-shaped probe), 8016(Hook type prod), 8212-USB(USB adapter) 1, 9186A(Carrying case), 9187(Cord case)						

3552/3552BT only

- *2 Low-resistance range is protected by a built-in fuse (0.5A/ 1000V, Dia. 6.3 x 32 mm)
- *3 3552BT only, Bluetooth® is a trademark or registered trademark of Bluetooth SIG, Inc. Some countries regulate the compliance with their Radio Law of the products equipped with Bluetooth®. Please confirm it with your distributor before purchasing our products equipped with Bluetooth®

The right model to meet your needs

Model	6 ranges insulation resistance	Memory	PC communi- cation	Bluetooth®
KEW 3551	•			
KEW 3552	•	•	•	
KEW 3552BT	•	•	•	•

Accessories



Optional accessories





Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely Safety Warnings: for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

For inquiries or orders:



KYORITSU ELECTRICAL **INSTRUMENTS** WORKS, LTD.

2-5-20, Nakane, Meguro-ku, Tokyo, 152-0031 Japan Phone:+81-3-3723-0131 Fax:+81-3-3723-0152

www.kew-ltd.co.jp

